Installation Instructions
Filter Intake Kit

Description:
Apollo Dental Products' Filter Intake Kits protect compressors from incoming dust and debris. They should be changed annually as part of a comprehensive maintenance schedule. The filter is included with necessary replacement hose to improve the location of incoming air.



Installation:

Step 1: Remove old filter assembly from intake port on compressor head.

Step 2: Replace filtercartridge with new cartridge provided.

Step 3: Insert hose into clamp by opening clamp with pliers.

Step 4: Insert filter into hose and close clamp.

Warranty Information: 90 Days

All ADP accessories are thoroughly inspected and tested in accordance with rigid specifications and standards. Our accessories are guaranteed against any defective material and workmanship from the date of shipment; provided that the installation, operation, and maintenance is done in accordance with ADP procedures as outlined in our Installation and Maintenance Guides. Warranty cards must be returned to ADP within ten days of installation to effect warranty. No other warranties or guarantees, expressed or implied are made.

ADP's obligation under the warranty is to provide parts for the repair or, at its option, to provide the replacement product (excluding labor). All special, incidental and/or consequential damages are excluded. We will not issue credit for accessories without first attempting to correct the problem in the field. Written notice of breach of warranty must be given to ADP within the warranty period. The warranty does not cove damage resulting form improper installation or maintenance, accident or misuse. The warranty does not cover damage resulting from the use of cleaning, disinfecting or sterilizing chemicals and processes. Failure to follow instructions provided in ADP's Installation and Maintenance Guides may void the warranty.

Apollo Dental Products, Inc. • 245 W. Dakota Ave. • Clovis, CA 93612 • Technical Support
559/292-1444 • 800/233-4151 • Fax 559/292-1555 • www.apollodental.com

AMI60732 Rev. 3/02